



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/623,793	09/08/2000	Wesley H Verkaart	70869-0068US	6666

7590 07/06/2004

Clark & Brody
1750 K Street N W Suite 600
Washington, DC 20006

EXAMINER

SIEFKE, SAMUEL P

ART UNIT	PAPER NUMBER
----------	--------------

1743

DATE MAILED: 07/06/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/623,793

Applicant(s)

VERKAART ET AL.

Examiner

Samuel P Siefke

Art Unit

1743

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3-10 and 14-19 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1,3-10 and 14-19 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. ____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date ____.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. ____.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: ____.

DETAILED ACTION

Response to Arguments

Applicant's arguments filed 2/20/04 have been fully considered but they are not persuasive. Applicant argues, "Esposito shows an enclosure for a syringe that allows a fluid to be injected into the syringe through the fitting member 50. Esposito does not show a flexible part that would allow the user to operate the plunger." The Office would like to point out that claim 1 recites limitations on the manner in which the syringe is used, "to allow operation of said plunger when said syringe is in said casing and said first and second parts are connected." Such limitations are not attributed patentable weight in claimed to a device. It is also noted that Esposito discloses the sheath may be made from a flexible material such as a thermoplastic material (col. 5, line 6, fig. 2, shows flexibility). Therefore all the Esposito includes all the limitations set forth by claim 1.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims **1,3-10** and **14-19** are rejected under 35 U.S.C. 102(e) as being anticipated by Esposito (USPN 5,997,811).

Esposito discloses a syringe and a syringe package sterile apparatus that comprises: a syringe for transferring non-sterile; a casing first part (50) and a second part (30; col. 4, lines 1-50) configured to receive a sterile syringe (20) containing a non-sterile fluid (col. 2, lines 4-9). The non-sterile fluid flows through the first part (50) and into the sterile syringe, which is in the sterile package. The first part (50, lap joint or luer connector, col. 2, lines 13-29) receives the barrel portion of the syringe and the second part (30) receives the plunger portion of syringe (see fig. 1). The first part (50) is a rigid tube and the second part (30) is flexible sheet (sheath, shape of a bag)(see fig. 1; col. 4, lines 1-12). The second part (30) creates a bellows when the user applies pressure to the plunger to draw fluid into the syringe. The sterile sheath of the sterile syringe package can be made of a radiation stable thermoplastic material or a breathable material sterilizable by gas or steam (col. 2, lines 24-28). It is noted that Applicant recites limitations on the manner in which the syringe is used, "to allow operation of said plunger when said syringe is in said casing and said first and second parts are connected." Such limitations are not attributed patentable weight in claimed to a device. It is also noted that Esposito discloses the sheath may be made from a flexible material such as a thermoplastic material (col. 5, line 6, fig. 2, shows flexibility).

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Samuel P Siefke whose telephone number is 571-272-1262. The examiner can normally be reached on M-F 7:00am-5:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on 571-272-1700. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.


Art Unit: 1743

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Sam P. Siefke

A handwritten signature in black ink, consisting of a stylized, overlapping loop structure.

June 28, 2004

A handwritten signature in black ink, appearing to read "Jill Warden".

Jill Warden
Supervisory Patent Examiner
Technology Center 1700